

# Patient MRI Form



**Neurology:** Please complete this form for your patient to take to the MRI center and make a copy to include in the patient medical file.

**Radiology:** Using the data from this form, input the model number and implant location in the Determining MRI Scan Conditions tool or refer to latest MRI Instructions For Use (IFU) available online through [www.VNSTherapy.com](http://www.VNSTherapy.com) for scan conditions before proceeding with an MRI.

## CONTACT INFORMATION

Patient Name: \_\_\_\_\_

Physician Name: \_\_\_\_\_

Physician Phone Number: \_\_\_\_\_

Clinic Name: \_\_\_\_\_

Clinic Address: \_\_\_\_\_

## GENERATOR INFORMATION

Generator Model # \_\_\_\_\_

Generator Serial # \_\_\_\_\_

Implant Location (VNS Therapy system located between C7 and T8 vertebrae)

Upper left chest at or above armpit (above rib 4)

Other Locations \_\_\_\_\_

## Perform the following to prepare the patient for an MRI:

Interrogate the VNS Therapy® generator and record the generator settings

When an interrogation is performed, the generator serial number, implant date, and stimulation parameters are logged in the programmer database and available to view.

Normal Output Current (mA)	
Signal Frequency (Hz)*	
Pulse Width (µsec)*	
Signal On Time (sec)*	
Signal Off Time (min)*	
Magnet Output Current (mA)	
Magnet On Time (sec)*	
Magnet Pulse Width (µsec)*	

\*Optional for Model 103 and newer devices

### Models 106 and 1000 Only

AutoStim Output Current (mA)	
Tachycardia Detection (On/Off)	

### Model 1000 Only

Low Heart Rate Detection (On/Off)	
Low Heart Rate Threshold (bpm, if on)	
Prone Position Detection (On/Off)	
Scheduled Programming (On/Off)	
Day-Night Programming (On/Off)	

Perform System Diagnostics to ensure proper operation of the generator.

Reprogram the VNS Therapy stimulation output to OFF:

\*\* For select models with AutoStim Mode

- Output current (mA): 0.0
- Magnet Current (mA): 0.0
- AutoStim Current\*\* (mA): 0.0 and Tachycardia\*\* detection to OFF

Turn off any other optional device features (Model 1000 only).

Interrogate the generator to verify that the programming was successful.

Determine if the placement of the VNS Therapy generator is between C7-T8 vertebrae and located in the upper left chest area, above rib 4.

Make an appointment for an office visit following the MRI procedure to program the VNS Therapy stimulation ON.

I have programmed the patient's VNS Therapy system OFF on \_\_\_\_\_ at \_\_\_\_\_ am/pm to facilitate their MRI scan at your facility.

Signature: \_\_\_\_\_

Print Name: \_\_\_\_\_

Have questions? Refer to [www.easy-mri.com](http://www.easy-mri.com) for MRI labeling and safety information OR contact LivaNova Clinical Technical Services at 1-866-882-8804

The device has been evaluated for MRI induced risks, including heating, unintended stimulation, force, torque, device malfunction and device vibration and has been determined to be safe under the conditions specified in labeling; however, the patient may feel sensations of warmth or vibration at the implant site during the MRI scan.

LivaNova USA, Inc.  
100 Cyberonics Boulevard  
Houston, Texas 77058  
Tel: +1.800.332.1375  
Fax: +1.281.218.9332  
[www.VNSTherapy.com](http://www.VNSTherapy.com)

LivaNova Belgium NV  
Ikaroslaan 83  
1930 Zaventem  
Belgium  
Tel: +32.2.720.95.93  
Fax: +32.2.720.60.53

©2020 LivaNova USA, Inc, a wholly-owned subsidiary of LivaNova PLC. All rights reserved.  
LivaNova®, AspireSR® and VNS Therapy® are registered trademarks of LivaNova USA, Inc.

**LivaNova**  
Health innovation that matters

PtMRIFrm20U11